



Deadline Summary

2nd Sight Solutions

ID	Deadline	Days left	Warnings
1	RSU: verify Reagent Additions	165	
2	RSU: verify RV Bubbler Function	162	
3	RSU: verify Elevator VD Reproducibility	193	
4	RSU: verify RV Elevator Position Feedback System	193	
5	RSU: verify Oil Bath Temperature Regulation	146	
6	GC: replace Gas Chromatograph Septum	22	
7	TLC: calibrate Background Activity Check	55	
8	TLC: calibrate TLC: calibrate Resolution Calibration	49	
9	TLC: calibrate Positional Linearity Check	56	
10	TLC: calibrate TLC: calibrate Normalization Calibration	56	
11	TLC: calibrate Detector Linearity	56	
12	Laminar Flow measure Air Flow	93	
13	Laminar Flow verify HEPA filter integrity	93	
14	Oven: verify Heat Distribution Study	63	
15	Oven: calibrate Thermocouple Calibration	180	
16	Lab: verify Environmental Microbial Testing	84	
17	Laminar Flow verify Microbial Testing		(missing data)
18	FDG: verify Sterility Testing of the FDG Product	23	
19	QC: calibrate Multi Channel Analyzer Calibration	48	
20	Capintec: calibrate General Calibration		(missing data)
21	Capintec: calibrate Linearity	167	
22	FDG: verify Isotonicity of the FDG Product Q022	109	
23	balance: calibrate Calibration of Precision Balance	1	Due soon - 1st warning.
24	LAL: prepare Control Standard Endotoxin Dilutions	8	
25	K222: prepare K222 TLC Strips	45	
26	K222: prepare K222 Strip Calibration Solutions	146	
27	TLC: prepare FDG QC TLC Solvent	19	
28	solutions: prepare I/E columns	19	
29	misc lab equip calibrate Stopwatch #1	516	
30	misc lab equip calibrate Stopwatch #2	63	
31	misc lab equip calibrate 1000 ul pipette	184	
32	misc lab equip calibrate 200 ul pipette	184	
33	TLC: measure Rf determination of new batch of plates	103	
34	misc lab equip calibrate Dry bath analog thermometer	1876	
35	RSU: calibrate Terminal gas flow meter	27	Due soon - 1st warning.
36	RSU: calibrate Dual thermometer Fisher, SN 51214227	424	
37	RSU: calibrate Temp controllers		(missing data)
38	RSU: calibrate Pressure		(missing data)

ID	Deadline		Days left	Warnings
39	GC:	prepare Standard soln	13	
40	lab:	Renew Nuclear license renewal	1345	
41	RSU:	replace Hot cell HEPA filter change		(missing data)
43	misc lab equip	verify Fridge temp		(missing data)
44	solutions:	filter Ethanol for C-18 conditioning	23	
45	misc lab equip	calibrate Fisher digital thermometer SN 20238389	63	
46	K222:	prepare Kryptofix solution	8	
47	RSU:	calibrate In-line N2 flow meter	179	
48	misc lab equip	calibrate Dry bath digital thermometer		(missing data)
50	Lab:	Renew CNSC Annual Compliance Report	268	
51	Laminar Flow	replace Laminar Flow Hood Pre-filter	56	
52	RSU:	verify Hot Cell Environmental Microbial Testing	84	
53	GC:	calibrate Residual solvent cal. curves		(missing data)
54	GC:	replace Capillary column		(missing data)
55	Capintec:	calibrate Geometric independence		(missing data)
56	LAL:	verify LAL label claim sensitivity	12	Due soon - 1st warning.
57	RSU:	verify Bioburden load of CPCU	29	Due soon - 1st warning.
58	GC:	Renew Column impurities bakeout	23	
59	GC:	calibrate Replicate injections		(missing data)
60	Lab:	clean Floors, walls and ceiling	56	
61	FDG:	prepare Serum reagent vial septa	24	
62	Lab:	verify RM 14 survey		(missing data)
63	Oven:	clean glassware depyrogenation	1	Due soon - 1st warning.
64	Laminar Flow	verify Media fill simulation	29	
65	RSU:	replace Pinch valve tubing and connectors	66	
66	RSU:	clean Reaction vessel cleaning	4	Due soon - 1st warning.
67	solutions:	prepare Iodoplatinate solution new	145	
68	LAL:	test New batch of CSE dilutions	10	
69	RSU:	replace Nitrogen sterilization filter	56	
70	misc lab equip	calibrate Temperature humidity meter		(missing data)